

BIOCOMPATIBILITY Certificate

Test Item: **SUCASLIDE**
carbon and chromium surface coating

Supplier: **Eifeler Werkzeuge GmbH, Duderstädter Straße 14, 40595 Düsseldorf**

Studies performed: The following studies were performed in order to determine the biocompatibility of the product according to ISO 10993-1:

CYTOTOXICITY (BSL Study No.: 112434)
INTRACUTANEOUS REACTIVITY (BSL Study No.: 112435)
SENSITISATION ¹⁾ (BSL Study No.: 112436)

1) LLNA (Local Lymph Node Assay)


Results: **The product did pass the requirements in the studies performed.
Therefore, the biocompatibility of the test material was proved.**

BSL BIOSERVICE Scientific Laboratories GmbH Munich

Behringstraße 6

82152 Planegg

Germany



Dr. Katharina Lütkenhaus

Biological Safety Testing

Date: 19 July 2011



BSL BIOSERVICE Scientific Laboratories GmbH

Behringstrasse 6/8 · 82152 Planegg, Germany
Telefon +49(0)89-899 65 00 · Fax +49(0)89-899 65 011
e-mail: info@bioservice.com · www.bioservice.com
Geschäftsführer: Dr. Wolfram Riedel
Amtsgericht München, HRB 109 770
Erfüllung und Gerichtsstand München
NORD/LB Norddeutsche Girozentrale, BLZ 250 500 00, Kto. 151 328 523, Swift-BIC: NOLADE2HXXX, IBAN: DE76 2505 0000 0151 3285 23
Deutsche Bank München, BLZ 700 700 24, Kto. 9 407 750, Swift-BIC: DEUTDE33MUC, IBAN: DE52 7007 0024 0940 7750 00



Akkreditiert durch
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln
und Medizinprodukten
ZLG-P-986.96.01